



Noncarbapenemase OXA-48 Variants (OXA-163 and OXA-405) Falsely Detected as Carbapenemases by the β Carba Test

 Laurent Dortet, Thierry Naas

Associated French National Reference Center for Antibiotic Resistance, Le Kremlin-Bicêtre, France; Research Unit EA7361 "Structure, Dynamic, Function and Expression of Broad Spectrum β -Lactamases," Faculty of Medicine, University Paris-Sud, Le Kremlin-Bicêtre, France; Department of Bacteriology-Parasitology-Hygiene, Bicêtre Hospital, Assistance Publique—Hôpitaux de Paris, Le Kremlin-Bicêtre, France; Joint Research Unit EERA "Evolution and Ecology of Resistance to Antibiotics," Institut Pasteur-APHP-University Paris Sud, Paris, France

KEYWORDS rapid diagnostics, carbapenemases, OXA-48, OXA-163, OXA-405, OXA-247, OXA-405

In *Enterobacteriaceae*, carbapenem resistance may be mediated by (i) plasmid-encoded or chromosome-encoded cephalosporinase and/or extended-spectrum β -lactamase production associated with decrease permeability of the outer membrane and (ii) carbapenemase production (1). During the last decade, the rapid dissemination of carbapenemase-producing *Enterobacteriaceae* (CPE) led to fear of a return to a preantibiotic era in which no antimicrobial molecule might be useful for the treatment of infected patients (1). To limit the spread of CPE, a lot of countries have implemented strict infection control measures for infected and colonized patients. However, implementation of these infection control measure is costly. Accordingly, there is an urgent need for accurate and fast diagnostic tests able to identify CPE among carbapenem-resistant isolates (2). For that purpose, several diagnostic tests have been developed during the last 5 years. They include (i) tests for carbapenemase inhibition activity (2); (ii) the carbapenem inactivation method (3); (iii) detection of carbapenem hydrolysis by matrix-assisted laser desorption ionization–time of flight mass spectrometry (4), biochemical tests (e.g., Carba NP test and derivatives) (5, 6), or an electrochemical method (BYG test) (7); (iv) immunochromatographic assays aiming to detect OXA-48-like CPE, IMP-like CPE, and OXA-48/KPC CPE (8–10); and (v) diverse molecular tools able to detect the most prevalent carbapenemase-encoding genes (11, 12).

Among the biochemical tests able to detect carbapenem hydrolysis activity, a new colorimetric assay, named the β Carba test (Bio-Rad, Marnes La Coquettes, France), based on the color change (yellow to red) of a chromogenic substrate in the presence of a carbapenemase has been recently commercialized. Recently, this test has been evaluated by Compain et al. on a collection of 79 enterobacterial isolates including 30 carbapenemase producers, 12 noncarbapenemase producers with decreased susceptibility to carbapenems, and 37 isolates susceptible to carbapenems (13). In that paper, it was claimed that the β Carba test was 100% specific and 100% sensitive. However, OXA-163-producing *Klebsiella pneumoniae* and OXA-405-producing *Serratia marcescens* gave positive results at 30 min, 1 h, and 3 h with the β Carba test, whereas they were not CPE. Indeed, the OXA-48 group is composed of enzymes with true carbapenemase activity (OXA-48, OXA-181, OXA-204, OXA-232, OXA-244, OXA-245, OXA-370, and OXA-484) and of enzymes devoid of any carbapenemase activity because of an amino acid deletion in their active site (OXA-163, OXA-247, and OXA-405) (14, 15). Because of their close relationship with OXA-48, these three noncarbapenemase variants usually gave false-positive results with molecular methods (11). Since Compain et al. used a molec-

Citation Dortet L, Naas T. 2017.

Noncarbapenemase OXA-48 variants (OXA-163 and OXA-405) falsely detected as carbapenemases by the β Carba test. *J Clin Microbiol* 55:654–655. <https://doi.org/10.1128/JCM.02086-16>.

Editor Alexander J. McAdam, Boston Children's Hospital

Copyright © 2017 American Society for Microbiology. All Rights Reserved.

Address correspondence to Laurent Dortet, laurent.dortet@bct.aphp.fr.

For the author reply, see <https://doi.org/10.1128/JCM.02114-16>.

ular method as a reference, OXA-163 and OXA-405 were falsely considered true carbapenemases. Actually, they are false positive by the β Carba test. Consequently, the performance of the β Carba test on the collection tested is 100% sensitivity and 96.1% specificity at 1 h and 85.7% sensitivity and 96.1% specificity at 30 min, which is the time recommended by the manufacturer. Of note, the other colorimetric tests able to detect carbapenemase activity, named the Carba NP test, the Rapidec Carba NP (bioMérieux, La Balmes les Grottes, France), and the Rapid Carba Screen (ROSCO Diagnostica, Taastrup, Denmark), were previously shown to perfectly detect OXA-163 and OXA-405 as noncarbapenemases (5).

REFERENCES

1. Nordmann P, Naas T, Poirel L. 2011. Global spread of carbapenemase-producing Enterobacteriaceae. *Emerg Infect Dis* 17:1791–1798. <https://doi.org/10.3201/eid1710.110655>.
2. Hrabák J, Chudackova E, Papagiannitsis CC. 2014. Detection of carbapenemases in Enterobacteriaceae: a challenge for diagnostic microbiological laboratories. *Clin Microbiol Infect* 20:839–853. <https://doi.org/10.1111/1469-0691.12678>.
3. van der Zwaluw K, de Haan A, Pluister GN, Bootsma HJ, de Neeling AJ, Schouls LM. 2015. The carbapenem inactivation method (CIM), a simple and low-cost alternative for the Carba NP test to assess phenotypic carbapenemase activity in Gram-negative rods. *PLoS One* 10:e0123690. <https://doi.org/10.1371/journal.pone.0123690>.
4. Lasserre C, De Saint Martin L, Cuzon G, Bogaerts P, Lamar E, Glupczynski Y, Naas T, Tande D. 2015. Efficient detection of carbapenemase activity in Enterobacteriaceae by matrix-assisted laser desorption/ionization–time of flight mass spectrometry in less than 30 minutes. *J Clin Microbiol* 53:2163–2171. <https://doi.org/10.1128/JCM.03467-14>.
5. Dortet L, Agathine A, Naas T, Cuzon G, Poirel L, Nordmann P. 2015. Evaluation of the RAPIDEC® CARBA NP, the Rapid CARB Screen® and the Carba NP test for biochemical detection of carbapenemase-producing Enterobacteriaceae. *J Antimicrob Chemother* 70:3014–3022. <https://doi.org/10.1093/jac/dkv213>.
6. Kabir MH, Meunier D, Hopkins KL, Giske CG, Woodford N. 2016. A two-centre evaluation of RAPIDEC® CARBA NP for carbapenemase detection in Enterobacteriaceae, *Pseudomonas aeruginosa* and *Acinetobacter* spp. *J Antimicrob Chemother* 71:1213–1216. <https://doi.org/10.1093/jac/dkv468>.
7. Bogaerts P, Yunus S, Massart M, Huang TD, Glupczynski Y. 2016. Evaluation of the BYG Carba test, a new electrochemical assay for rapid laboratory detection of carbapenemase-producing Enterobacteriaceae. *J Clin Microbiol* 54:349–358. <https://doi.org/10.1128/JCM.02404-15>.
8. Dortet L, Jousset A, Sainte-Rose V, Cuzon G, Naas T. 2016. Prospective evaluation of the OXA-48 K-SeT assay, an immunochromatographic test for the rapid detection of OXA-48-type carbapenemases. *J Antimicrob Chemother* 71:1834–1840. <https://doi.org/10.1093/jac/dkw058>.
9. Glupczynski Y, Evrard S, Ote I, Mertens P, Huang TD, Leclipteux T, Bogaerts P. 2016. Evaluation of two new commercial immunochromatographic assays for the rapid detection of OXA-48 and KPC carbapenemases from cultured bacteria. *J Antimicrob Chemother* 71:1217–1222. <https://doi.org/10.1093/jac/dkv472>.
10. Kitao T, Miyoshi-Akiyama T, Tanaka M, Narahara K, Shimajima M, Kirikae T. 2011. Development of an immunochromatographic assay for diagnosing the production of IMP-type metallo- β -lactamases that mediate carbapenem resistance in *Pseudomonas*. *J Microbiol Methods* 87:330–337. <https://doi.org/10.1016/j.mimet.2011.09.011>.
11. Dortet L, Fusaro M, Naas T. 2016. Improvement of the Xpert Carba-R kit for the detection of carbapenemase-producing Enterobacteriaceae. *Antimicrob Agents Chemother* 60:3832–3837. <https://doi.org/10.1128/AAC.00517-16>.
12. Findlay J, Hopkins KL, Meunier D, Woodford N. 2015. Evaluation of three commercial assays for rapid detection of genes encoding clinically relevant carbapenemases in cultured bacteria. *J Antimicrob Chemother* 70:1338–1342. <https://doi.org/10.1093/jac/dku571>.
13. Compain F, Gallahd S, Eckert C, Arlet G, Ramahefasolo A, Decre D, Lavollay M, Podglajen I. 2016. Assessment of carbapenem resistance in Enterobacteriaceae with the rapid and easy-to-use chromogenic β Carba test. *J Clin Microbiol* 54:3065–3068. <https://doi.org/10.1128/JCM.01912-16>.
14. Dortet L, Oueslati S, Jeannot K, Tande D, Naas T, Nordmann P. 2015. Genetic and biochemical characterization of OXA-405, an OXA-48-type extended-spectrum β -lactamase without significant carbapenemase activity. *Antimicrob Agents Chemother* 59:3823–3828. <https://doi.org/10.1128/AAC.05058-14>.
15. Oueslati S, Nordmann P, Poirel L. 2015. Heterogeneous hydrolytic features for OXA-48-like β -lactamases. *J Antimicrob Chemother* 70:1059–1063. <https://doi.org/10.1093/jac/dku524>.